

VERIFICATION OF CONFORMITY

Applicant Name & Address : Shanghai Yihua V&A Instrument Co., Ltd.
881 yecheng road, jiading district, shanghai, china

Product(s) : multimeter

Model(s) : VA19

Technical Specification : Rated input: 9 V DC (6F22 battery)
Measuring circuits:
Rated voltage to earth & rated voltage: 600 V AC or DC
Rated current: 400 mA AC or DC, or 10 A AC or DC
Resistance: 40 M Ω ;
Measurement category: CAT III

Brand name : V&A

Relevant Standard(s) : EN 61010-1:2010
EN 61010-2-033:2012
EN 61010-031:2015

Directive(s) : Low Voltage Directive 2014/35/EU

Verification Number : EFSH201708-L031

Report Number(s) : EFSH17070078-IE-01-L01

NOTE 1: This verification is part of the full test report(s) and should be read in conjunction with it.


This is the result of tests carried out on those samples of the product referred to above which were submitted for testing, in accordance with the specification for the respective standards. The sample(s) of the tested product has been found to comply with the relevant standard/specification to the directive(s) listed on this verification at the time the tests were carried out.

The manufacturer according to definition by EU directive may indicate compliance to said directive(s) by signing a DoC himself and applying the CE-marking to products identical to the tested sample(s). In addition, the manufacturer shall file and keep the documentation according to the rules of the applicable directive(s) and shall consider changes of the standard(s) if relevant. Additional requirements may be applicable such as additional directives or local laws.

The 'CE' marking shall consist in the initials 'CE' taking the following form:



The 'CE' marking must be affixed to the product/s or to its marking plate. Where this is not possible or not warranted on account of the nature of the products, it must be affixed to the packaging, if any, and to the accompanying documents



Jack Gan
Supervisor
2017-08-23