

VERIFICATION OF CONFORMITY

Applicant Name & Address : Shanghai Yihua V&A Instrument Co., Ltd.
881 Yecheng Road, Jiading District, Shanghai, China.

Product(s) : multimeter

Model(s) : VA19

Technical Specification : Rated input: 9VDC (6F22 battery)
Measuring circuits:
Rated voltage to earth & rated voltage: 600V AC or DC
Rated current: 400mA AC or DC or 10A AC or DC
Resistance: 40 M
Measurement category: CAT III

Brand name : V&A

Relevant Standard(s) : EN 61326-1 :2013
EN 61326-2-2 :2013

Directive(s) : EMC Directive 2014/30/EU

Verification Number : EFSH201708-E081

Report Number(s) : EFSH17070078-IE-02-E01

Project Number : EFSH17070078-IE-02

NOTE 1: This verification is part of the full test report(s) and should be read in conjunction with it.

This is the result of tests carried out on those samples of the product referred to above which were submitted for testing, in accordance with the specification for the respective standards. The sample(s) of the tested product has been found to comply with the relevant standard/specification to the directive(s) listed on this verification at the time the tests were carried out.

The manufacturer according to definition by EU directive may indicate compliance to said directive(s) by signing a DoC himself and applying the CE-marking to products identical to the tested sample(s). In addition, the manufacturer shall file and keep the documentation according to the rules of the applicable directive(s) and shall consider changes of the standard(s) if relevant. Additional requirements may be applicable such as additional directives or local laws.

The 'CE' marking shall consist in the initials 'CE' taking the following form:



The 'CE' marking must be affixed to the product/s or to its marking plate. Where this is not possible or not warranted on account of the nature of the products, it must be affixed to the packaging, if any, and to the accompanying documents.



Dio Yang
Manager
2017-08-28